




# PRECISION DIGITAL DEPTH GAUGE MICROTECH



ISO/IES 17025:2006 (cert. № 4K051)



Quality management system  
ISO 9001:2015 (cert. № U228396)





# 1.MODIFICATIONS

Item No	Range	Resolution	Accuracy	Base	Waterproof	Data output
	mm	mm	mm	mm		
1432010151	0-150	0,005	±0,03	102	IP-54	RS-232
1432010201	0-200	0,005	±0,03	102	IP-54	RS-232
1432010301	0-300	0,005	±0,04	102	IP-54	RS-232

## 2.BUTTON FUNCTIONS

Switch between metric  
and English measurement  
systems

Switching on/off



Setting base ZERO point  
(5 sec switch)

Switching between absolute  
and relative measurement  
(display INC) or setting ZERO





## 3. OPERATION INSTRUCTIONS

3.1 Wipe with a clean cloth, soaked in gasoline, measuring surface of the frame and gauge calipers to remove anti-corrosion oil. Then wipe them with a clean dry cloth.

3.2 If necessary, open the battery cover; insert the battery (type CR1632) according to the polarity of the electrodes. Blinking display information or absence suggests replacing battery.

3.3 This depth gauge has Autoswitch on/ off function:

- move electronic module for switch on depth gauge
- after 10 minutes without any moving caliper will switch off

3.4 During the measurement avoid warps of measuring surfaces of the instrument. Measuring surface must be fully in contact with the measurement object.



## WARNING!

**IN THE PROCESS OF WORKING WITH DEPTH GAUGE SHOULD BE AVOIDED:**

**Scratches on the measuring surfaces;**

**Measuring the size of object in the process of machining;**

**Shocks or dropping, avoid bending of rod or other surfaces.**





## 4. INSTRUMENT FIGURE

