




DIAL THICKNESS GAUGE MICROTECH



ISO/IES 17025:2006 (cert. № 4K051)



Quality management system
ISO 9001:2015 (cert. № U228396)





1.MODIFICATIONS

Item No	Range	Depth	Resolution	Accuracy	Inidcation
	mm	mm	mm	mm	
137001020	0-1	25	0,001	±0,005	Dial
137010120	0-10	120	0,01	±0,030	Dial
137010200	0-10	200	0,01	±0,030	Dial
137020120	0-20	120	0,01	±0,040	Dial
137025200	0-25	200	0,01	±0,040	Dial
137050100	0-50	100	0,01	±0,040	Dial
137050200	0-50	200	0,01	±0,050	Dial
137100100	0-100	100	0,01	±0,040	Dial





2. OPERATION INSTRUCTIONS

- 2.1 Clean the oil from the measuring rod and measuring anvil.
- 2.2 Check the zero setting.
- 2.3 Check the stability of the zero position of the 4-5x caging.
- 2.4 Wipe the testing object with a clean soft cloth.



WARNING !

Measuring rod should move without shock at the end of turn.

Avoid contact with the indicator and oil emulsion.

Do not turn indicator when it is fixed in the tool holder by a sleeve





3. INSTRUMENT FIGURE

